

Exhibit B

Sec. 96.12. AUTOMATED EXTERNAL DEFIBRILLATOR AND STOP THE BLEED KITS.

(A) Purpose. Effective June 18, 2024, all new and existing buildings within the City limits of the City of Delray Beach shall provide Automated External Defibrillator devices and Bleeding Control Kit as required herein.

(B) Definitions.

Automated External Defibrillator (AED). A medical device designed to analyze the heart rhythm and deliver an electric shock to victims of ventricular fibrillation to restore a heart rhythm to normal.

Bleeding Control Kit. A specialized first aid kit designed to stop severe, life-threatening blood loss from traumatic injuries.

(C) Automated External Defibrillators and Bleeding Control Kits shall be installed in the following buildings:

- (1) Fitness centers, gymnasiums, and indoor recreational centers in excess of one thousand five hundred (1,500) square feet;
- (2) Theaters, restaurants, drinking establishments, with a capacity of one hundred (100) or greater;
- (3) Places of worship with a capacity of one hundred (100) or greater;
- (4) Commercial, Retail, Office buildings, and any business occupancies with a square footage greater than twenty thousand (20,000) square feet;
- (5) All dental offices in accordance with Florida Administrative Code 64B5-17.015;
- (6) All adult day care facilities;
- (7) All charter and private schools;
- (8) Assisted living facilities as defined by F.S. § 429.021(5) as amended from time to time;
- (9) Ambulatory surgical centers as defined by F.S. § 395.002(3) as amended from time to time;
- (10) Walk in medical care facilities;
- (11) Hospitals providing emergency services, including freestanding facilities, shall be excluded;
- (12) All hotels and motels; and
- (13) Multi-story residential/dormitory buildings five (5) floors or more.

(D) Multi-story occupancies listed in and meeting the requirements of subsection B shall place an AED device on every other floor.

(1) Residential elevators supplying services to an independent dwelling unit, only, shall be exempt from this requirement.

(E) Buildings that contain more than one AED device shall be required to maintain a Bleeding Control Kit with each defibrillator device.

(F) Installation and Operation.

(1) The City's Fire Marshal or designee shall inspect all AED devices and Bleeding Control Kits prior to being placed in service or available for use, and on an annual basis thereafter.

(2) AED devices shall be:

i. Conspicuously located in plain view of the primary public entrance, with unobstructed access;

ii. Housed in a cabinet with a clear window in the door, an audible alarm signaling the opening or a door, permanently affixed to a wall no more than forty-eight (48) inches above the floor;

iii. Located below a sign having a minimum area of seventy (70) square inches and containing the letters "AED" with the universally recognizable symbol for AED which shall be placed no more than sixty (60) inches above the floor; and

iv. or as determined by the Fire Marshal or designee.

(3) Bleeding Control Kits shall be located in the same cabinet as an AED device. If not feasible, then Bleeding Control Kits shall be placed in a location approved by the Fire Marshal or designee.

(4) AED devices shall contain adult and pediatric pads.

(5) All AED devices shall be used in accordance with the manufacturer's guidelines.

(6) It shall be the responsibility of the owner of the building to provide all necessary training for appropriate use of AED devices and Bleeding Control Kits and maintain AED devices and Bleeding Control Kits in accordance with manufacturer's recommended maintenance requirements and as required herein.

(7) If an AED device or Bleeding Control Kit is removed for repair, a replacement shall be provided if removed for a period exceeding 10 days.

(8) Penalties. Violations of this Section shall be subject to the penalties set forth in Section 10.99 of the Code of Ordinances, Chapter 3 of the Code of Ordinances, F.S. ch. 162, and/or any other means of enforcement available.